

BLANK PAGE






Indian Standard

SPECIFICATION FOR IMPACTOR FOR HIP PROSTHESIS

- 1. Scope Covers dimensional and other requirements for impactor used for hip prosthesis in orthopaedic surgery.
- 2. Shape and Dimensions As shown in Fig. 1.



Atl dimensions in millimetres.

FIG. 1 IMPACTOR FOR HIP PROSTHESIS

- 2.1 Deviation of ± 2.5 percent shall be allowed on all dimensions.
- 3. Material Nylon or aluminium alloy with nylon facing at the working end shall be used. The aluminium alloy shall conform to Designation NF3 of IS: 734-1967 'Specification for wrought aluminium and aluminium alloys, forgings stock and forgings (for general engineering purposes) (first revision). The nylon may be bonded to the impactor by screws of the same material as that of the impactor.
- 4. Workmanship and Finish The impactor shall be free from cracks, seams, burrs and other surface defects.
- 5. Marking The impactor shall be clearly and indelibly marked with the manufacturer's name, initials or recognized trade-mark.
- 5.1 /S/ Certification Marking Details available with the Indian Standards Institution.
- 6. Packing Shall be as agreed to between the purchaser and the supplier.

Adopted 15 October, 1976

@ December 1975, ISI

PROPER Cr